

DISPOSABLE MEDICAL SHOE COVER

- ANTI-PENETRATION OF ALCOHOL
- ANTI BLOOD SPLASH
- ANTI-BACTERIAL PARTICLES
- WATERPROOF



PROFESSIONAL TESTING

National Supervision and Inspection Center (Zhejiang) of Textile and Garments Quality
Ningbo Fibre Inspection Institute

TEST REPORT

Report No.: NBSW20200902 Item: Blue/S/1/2020
 Sample name: medical isolation shoe covers Sample received date: 8/12/2020
 Applicant: Ningbo Snowwolf Outdoor Apparel Co., Ltd
 Sample Qty.: 2 pair + one piece
 Test items: Listed as follows:

Test item	Test method	Standard Requirement	Result	Pass/Fail
Determination of formaldehyde, mg/kg		≤200	≤20	Pass
pH Value	GB 18401-2010	4.0-9.0	6.0	Pass
Fluorine content	GB/T 3923.1-2013	none	none	Pass
Breaking strength, N	GB/T 4744-2013	Wale: ≥65 Cusper: ≥65	110 96	Pass
Water resistance, kPa (after washing)	Standard specified by applicant GB 19842-2009	≥1.67	25.13	Pass
Spray rating, grade (before washing)		≥3	3-4	Pass

Remarks: Breaking strength, Water resistance and Spray rating are judged according to the relevant indicators of the standard GB 19842-2009

sample:

Approver:

Add. Building: Pz No. 1008, Jiangnan Road, Hi-tech Zone, Ningbo, China Postcode: 315048
 Tel: 86-574-85531968 http://www.nfibii.com.cn E-mail: nfi@nfiib.com.cn
 Fax: 86-574-85531920

Expected usage : general isolation in medical institutions, wards, test rooms, etc.